

5.1

ENGLISH

1. Obstructions Form in the Chest Tube

6.1

ENGLISH

1. Chest Tube Free Of Obstructions
2. Proximal End
3. When Placed, Secure Striple Guide To Proximal End Area "Click" Into Position
4. Striple Guide Help Intubation
5. Drainage Material Flows To Drainage Canister

- Do not reuse
- Use by
- Lot
- Barcode
- Sterilization using ethylene oxide
- Do not reinsert
- Manufacturer
- Consult instructions for Use
- REF Catalog number
- MSD Urinate
- Contacts
- Do not use if package is damaged
- CAUTION:** PMS® (U.S.A.) has notified this device to use by or on the order of a physician.
- Rx ONLY**

The PleuraFlow® System with FlowGuide®
For Single Use Only. Made in the U.S.A.

PLEURAFLOW
ACTIVE CLEARANCE TECHNOLOGY

ENGLISH The PleuraFlow® System with FlowGuide®
For Single Use Only. Made in the U.S.A.

Manufactured for:
ClearFlow, Inc.,
140 Technology, Suite 100
Irvine, CA 92618, USA
US Toll Free: (844) CLR-FLOW (257-3569)
Support@clearflow.com
www.clearflow.com

Rx ONLY

PMS Rev. A
Page 1 of 3

1.1

ENGLISH

1. Pleuraflex
2. Pleuraflex Chest Tube
3. Distal
4. Pleuraflex End Connects To Pleuraflex Chest Tube
5. Pleuraflex Cap to Tube
6. Close One away from Patient

2.1

ENGLISH

1. Must Cut On Cut Line

3.1

ENGLISH

1. Pleuraflex Chest Tube
2. Pleuraflex End
3. Pleuraflex Clearance Apparatus
4. Distal End
5. Drainage Canister
6. Drainage Tubing

4.1

ENGLISH

1. Clearance Loop is Advanced to Top Tip when not in Use
2. Striple Guide is in the Pleural Pocket

